



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Gen Lab*, using *Sebia Capillarys 3 Tera* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Capillarys 3 Tera	Calibrator Lot: 7092, 8092	Application: %NGSP=0.09148 x [IFCC] (mmol/mol) + 2.152
Reagent Lot: 27074, 21054, 23094	Calibrator Assigned Values: 35.5 mmol/mol, 84.4 mmol/mol	

Date of Certification: February 1, 2025

Certification Expires: February 1, 2026

A handwritten signature in black ink, appearing to read 'David Sacks', is written over a horizontal line.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Randie R Little PhD', is written over a horizontal line.

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read 'Sharon Amalf', is written over a horizontal line.

SRL director/ supervisor